

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).



2.0 PRODUCT DESCRIPTION

Product Code: Single Analyte Custom Grade Solution
Catalog Number: CGHF1
Lot Number: P2-HF683293
Matrix: tr. HNO₃
tr. HF
Value / Analyte(s): 1 000 µg/mL ea:
Hafnium
Starting Material: HfO₂
Starting Material Lot#: 1955
Starting Material Purity: 99.9867%

3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Value: 1001 ± 7 µg/mL
Density: 1.005 g/mL (measured at 20 ± 4 °C)
Assay Information:

Assay Method #1 **1001 ± 4 µg/mL**
ICP Assay NIST SRM 3122 Lot Number: 151120

- The Calculated Value is a value calculated from the weight of a starting material that has been certified directly vs. a National Institute of Standards and Technology (NIST) SRM/RM. See Sec 4.2 for balance traceability.

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$$X_{CRM/RM} = \sum(w_i)(X_i)$$

X_i = mean of Assay Method i with standard uncertainty $u_{char i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$$w_i = (1/u_{char i}^2) / (\sum(1/u_{char i}^2))$$

$$CRM/RM \text{ Expanded Uncertainty } (\pm) = U_{CRM/RM} = k (u_{char}^2 + u_{bb}^2 + u_{lts}^2 + u_{ts}^2)^{1/2}$$

k = coverage factor = 2

$u_{char} = [\sum(w_i)^2 (u_{char i})^2]^{1/2}$ where $u_{char i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$$X_{CRM/RM} = (X_a)(u_{char a})$$

X_a = mean of Assay Method A with

$u_{char a}$ = the standard uncertainty of characterization Method A

$$CRM/RM \text{ Expanded Uncertainty } (\pm) = U_{CRM/RM} = k (u_{char a}^2 + u_{bb}^2 + u_{lts}^2 + u_{ts}^2)^{1/2}$$

k = coverage factor = 2

$u_{char a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

CRM/RMs are tested for trace metallic impurities by Axial ICP-OES and ICP-MS. The result from the most sensitive method for each element, is reported below. Solutions tested by ICP-MS were analyzed in an ULPA-Filtered Clean Room. An ULPA-Filter is 99.9985% efficient for the removal of particles down to 0.3 µm.

M	Ag	<	0.000578	M	Eu	<	0.000289	O	Na		0.001135	M	Se	<	0.027753	O	Zn		0.000166
M	Al	<	0.005203	M	Fe	<	0.232434	M	Nb		0.000447	O	Si		0.015743	O	Zr		0.013515
M	As	<	0.008672	M	Ga	<	0.000289	M	Nd	<	0.000289	M	Sm	<	0.000289				
O	Au	<	0.064460	M	Gd	<	0.000289	M	Ni	<	0.007516	M	Sn	<	0.002890				
M	B	<	0.007227	M	Ge	<	0.001156	M	Os	<	0.000288	O	Sr	<	0.000703				
M	Ba		0.000082	s	Hf	<		i	P	<		O	Ta	<	0.117200				
M	Be	<	0.002890	M	Hg	<	0.000288	M	Pb		0.000541	M	Tb	<	0.000289				
i	Bi	<		M	Ho	<	0.000289	M	Pd	<	0.002312	M	Te	<	0.000578				
M	Ca	<	0.099738	M	In	<	0.000289	M	Pr	<	0.000289	M	Th	<	0.004047				
M	Cd	<	0.000289	i	Ir	<		O	Pt	<	0.117200	M	Ti		0.001776				
M	Ce		0.000047	O	K	<	0.044536	M	Rb	<	0.001734	M	Tl		0.000247				
M	Co	<	0.000289	M	La	<	0.002312	M	Re	<	0.000289	M	Tm	<	0.000289				
M	Cr	<	0.016189	M	Li	<	0.001734	M	Rh	<	0.000289	M	U	<	0.002312				
M	Cs		0.000023	M	Lu		0.000058	M	Ru	<	0.000864	M	V	<	0.002312				
M	Cu		0.000376	M	Mg	<	0.006938	O	S		0.147464	M	W		0.000094				
M	Dy	<	0.000289	M	Mn	<	0.000867	M	Sb	<	0.000289	O	Y		0.002013				
M	Er	<	0.000289	M	Mo	<	0.003180	O	Sc	<	0.001523	M	Yb	<	0.000289				

M - Checked by ICP-MS O - Checked by ICP-OES i - Spectral Interference
n - Not Checked For s - Solution Standard Element

6.0 INTENDED USE

- For the calibration of analytical instruments and validation of analytical methods as appropriate.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit www.inorganicventures.com/TCT

Atomic Weight; Valence; Coordination Number; Chemical Form in Solution - 178.49 +4 6,7,8 Hf(F)6-2

Chemical Compatibility - Soluble in concentrated HCl, HF, H2SO4 (very hot) and HNO3. Avoid H3PO4 and neutral to basic media. Unstable at ppm levels with metals that would pull F- away (i.e. Do not mix with Alkaline or Rare Earths or high levels of transition elements unless they are fluorinated). Stable with most inorganic anions but precipitation with phosphate, oxalate, and tartrate with a tendency to hydrolyze forming the hydrated oxide in all dilute acids except HF.

Stability - 2-100 ppb levels stable alone or mixed with all other metals that are at comparable levels as the Hf(F)6-2+ Hf(OH)4F2-2 for months in 1% HNO3 / LDPE container. 1-10,000 ppm single element solutions as the Hf(F)6-2 chemically stable for years in 2-5% HNO3 / trace HF in a LDPE container.

Hf Containing Samples (Preparation and Solution) - Metal (Soluble in H2O / HF / HNO3); Oxide unlike TiO2 the HfO2 is best fused in one of the following ways (Na2O2 in NiO, Na2CO3 in Pt0 or Borax in Pt0); Organic Matrices (dry ash at 450 EC in Pt0 and dissolve by fusing with Na2CO3 and dissolving in HF / HNO3 / H2O).

Atomic Spectroscopic Information (ICP-OES D.L.s are given as radial/axial view):

Technique/Line	Estimated D.L.	Order	Interferences (underlined indicates severe)
ICP-MS 177amu	4 ppt	n/a	161Dy16O (fewer potential interferences on the 177 vs 180 line)
ICP-OES 264.141 nm	0.02 / 0.002 µg/mL	1	Ba, Th, U
ICP-OES 273.876 nm	0.02 / 0.002 µg/mL	1	U, Mo
ICP-OES 277.336 nm	0.02 / 0.002 µg/mL	1	Nb, Cr, U

HF Note: This standard should not be prepared or stored in glass.

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

September 28, 2019

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- **September 28, 2023**

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Approved By:

Michael Booth
Manager, Quality Control



Certifying Officer:

Paul Gaines
CEO, Senior Technical Director

