

300 Technology Drive  
Christiansburg, VA 24073 USA  
inorganicventures.com

P: 800-669-6799/540-585-3030  
F: 540-585-3012  
info@inorganicventures.com

## 1.0 ACCREDITATION / REGISTRATION

**INORGANIC VENTURES** is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (Amtivo Certificate Number 274357).



## 2.0 PRODUCT DESCRIPTION

Product Code: Single Analyte Custom Grade Solution  
Catalog Number: CGMSA1  
Lot Number: X2-S762837  
Matrix: H2O  
Value / Analyte(s): 1 000 µg/mL ea:  
Sulfur  
Starting Material: Methane Sulfonic Acid  
Starting Material Lot#: 2753  
Starting Material Purity: 100.0000%

## 3.0 CERTIFIED VALUES AND UNCERTAINTIES

**Certified Value:** 998 ± 4 µg/mL  
**Density:** 1.000 g/mL (measured at 20 ± 5 °C)

### Assay Information:

<b>Assay Method #1</b>	<b>996 ± 4 µg/mL</b> ICP Assay NIST SRM Traceable to 3154 Lot Number: T2-S727797
<b>Assay Method #2</b>	<b>998 ± 3 µg/mL</b> Acidimetric NIST SRM 84L Lot Number: 84L
<b>Assay Method #3</b>	<b>1000 ± 3 µg/mL</b> Calculated NIST SRM Lot Number: See Sec. 4.2

- The Calculated Value is a value calculated from the weight of a starting material that has been certified directly vs. a National Institute of Standards and Technology (NIST) SRM/RM. See Sec 4.2 for balance traceability.

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

### Characterization of CRM/RM by Two or More Methods

Certified Value,  $X_{CRM/RM}$ , where two or more methods of characterization are used is the weighted mean of the results:

$$X_{CRM/RM} = \sum(w_i)(X_i)$$

$X_i$  = mean of Assay Method  $i$  with standard uncertainty  $u_{char i}$

$w_i$  = the weighting factors for each method calculated using the inverse square of the variance:

$$w_i = (1/u_{char i}^2) / (\sum(1/u_{char i}^2))$$

$$CRM/RM \text{ Expanded Uncertainty } (\pm) = U_{CRM/RM} = k (u_{char}^2 + u_{bb}^2 + u_{lts}^2 + u_{ts}^2)^{1/2}$$

$k$  = coverage factor = 2

$u_{char} = [\sum(w_i)^2 (u_{char i})^2]^{1/2}$  where  $u_{char i}$  are the errors from each characterization method

$u_{bb}$  = bottle to bottle homogeneity standard uncertainty

$u_{lts}$  = long term stability standard uncertainty (storage)

$u_{ts}$  = transport stability standard uncertainty

### Characterization of CRM/RM by One Method

Certified Value,  $X_{CRM/RM}$ , where one method of characterization is used is the mean of individual results:

$$X_{CRM/RM} = (X_a) (u_{char a})$$

$X_a$  = mean of Assay Method A with

$u_{char a}$  = the standard uncertainty of characterization Method A

$$CRM/RM \text{ Expanded Uncertainty } (\pm) = U_{CRM/RM} = k (u_{char a}^2 + u_{bb}^2 + u_{lts}^2 + u_{ts}^2)^{1/2}$$

$k$  = coverage factor = 2

$u_{char a}$  = the errors from characterization

$u_{bb}$  = bottle to bottle homogeneity standard uncertainty

$u_{lts}$  = long term stability standard uncertainty (storage)

$u_{ts}$  = transport stability standard uncertainty

## 4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

### 4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

### 4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

### 4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

## 5.0 TRACE METALLIC IMPURITIES (TMI ) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

CRM/RMs are tested for trace metallic impurities by Axial ICP-OES and ICP-MS. The result from the most sensitive method for each element, is reported below. Solutions tested by ICP-MS were analyzed in an ULPA-Filtered Clean Room. An ULPA-Filter is 99.9985% efficient for the removal of particles down to 0.3 µm.

M Ag <	0.000630	M Eu <	0.000630	O Na	0.000052	M Se <	0.022000	O Zn <	0.001900
O Al <	0.006600	O Fe <	0.004200	M Nb <	0.000630	O Si <	0.021000	M Zr <	0.000630
M As <	0.004400	M Ga <	0.000630	M Nd <	0.000630	M Sm <	0.000630		
M Au <	0.002500	M Gd <	0.000630	O Ni <	0.005500	M Sn <	0.001300		
O B <	0.011000	M Ge <	0.001900	M Os <	0.000630	M Sr <	0.001300		
M Ba <	0.000630	M Hf <	0.000630	O P <	0.022000	M Ta <	0.000630		
O Be <	0.000440	M Hg <	0.005000	M Pb <	0.000630	M Tb <	0.000630		
M Bi <	0.000630	M Ho <	0.000630	M Pd <	0.001300	M Te <	0.001300		
O Ca	0.000071	M In <	0.000630	M Pr <	0.000630	M Th <	0.000630		
M Cd <	0.000630	M Ir <	0.000630	M Pt <	0.000630	M Ti <	0.003200		
M Ce <	0.000630	O K <	0.003600	M Rb <	0.000630	M Tl <	0.000630		
M Co <	0.000630	M La <	0.000630	M Re <	0.000630	M Tm <	0.000630		
O Cr <	0.006700	O Li <	0.000220	M Rh <	0.000630	M U <	0.000630		
M Cs <	0.001300	M Lu <	0.000630	M Ru <	0.000700	O V <	0.002200		
M Cu <	0.002500	O Mg <	0.000440	s S <		M W <	0.000940		
M Dy <	0.000630	M Mn <	0.001300	M Sb <	0.000630	M Y <	0.000630		
M Er <	0.000630	M Mo <	0.000630	M Sc <	0.002500	M Yb <	0.000630		

M - Checked by ICP-MS      O - Checked by ICP-OES      i - Spectral Interference  
n - Not Checked For      s - Solution Standard Element

## 6.0 INTENDED USE

**6.1** This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

**6.2** For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures [Terms and Conditions of Sale](https://www.inorganicventures.com/terms-and-conditions-sale). <https://www.inorganicventures.com/terms-and-conditions-sale>. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

## 7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

### 7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.
- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.
- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 25° C to minimize the effects of transpiration. Use at 20° ± 5° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit [www.inorganicventures.com/TCT](http://www.inorganicventures.com/TCT)

**Atomic Weight; Valence; Coordination Number; Chemical Form in Solution** - 32.07 +6 6 (O=)2 S(OH)2

**Chemical Compatibility** -Soluble in HCl, HNO3, H3PO4 and HF aqueous matrices water and NH4OH . Stable with all metals and inorganic anions at low to moderate ppm levels under acidic conditions except Ba and Pb and to a lesser extent Sr, and Ca.

**Stability** - 2-100 ppb levels- stability unknown- in 1% HNO3 / LDPE container. 1-10,000 ppm solutions chemically stable for years in LDPE container.

**S Containing Samples (Preparation and Solution)** -We most often get questions about the determination of S in Rocks, Silicates and insoluble sulfates (the finely powered sample is fused in a PtO crucible with 6 times its weight of Na2CO3 + 0.5 grams KNO3. The fuseate is extracted with water. Any BaSO4 present in the sample is transposed by the carbonate fusion to the BaCO3 which is left behind in the water-insoluble residue. If PbSO4 is present the fuseate should be boiled with a sodium carbonate saturated with CO2 solution for 1 hour or more where the PbSO4 will be transposed to the water insoluble carbonate which can be filtered off. Boiling the fuseate with a saturated carbonate solution is good insurance for samples containing Ba, Sr, and Ca. The Ba, Pb, Sr, Ca, free filtrate can be acidified and measured by ICP.)

**Atomic Spectroscopic Information (ICP-OES D.L.s are given as radial/axial view):**

Technique/Line	Estimated D.L.	Order	Interferences (underlined indicates severe)
ICP-MS 32 amu	30,000 ppt	n/a	16O2, 14N18O, 15N17O, 14N17O1H, 15N16O1H
ICP-OES 143.328 nm	0.4 / 0.035 µg/mL	1	
ICP-OES 166.669nm	0.2 / 0.19 µg/mL	1	Si, B
ICP-OES 182.034 nm	0.3 / 0.024 µg/mL	1	

## 8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

## 9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

**10.0 QUALITY STANDARD DOCUMENTATION**

**10.1 ISO 9001 Quality Management System Registration**

- Amtivo Certificate Number 274357

**10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"**

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

**10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"**

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

**11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY**

**11.1 Certification Issue Date**

June 22, 2026

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

**11.2 Lot Expiration Date**

- **June 22, 2031**

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

**11.3 Period of Validity**

- Sealed TCT Bag Open Date: \_\_\_\_\_

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

**12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS**

**Certificate Prepared By:**

Justin Dirico  
Stock Processing Supervisor



**Certificate Approved By:**

Shalin Presgraves  
SVS Coordinator



**Certifying Officer:**

Paul Gaines  
Chairman / Senior Technical Director

