

Certificate of Analysis

300 Technology Drive Christiansburg, VA 24073 USA inorganicventures.com

P: 800-669-6799/540-585-3030 F: 540-585-3012 info@inorganicventures.com

1.0 **ACCREDITATION / REGISTRATION**

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).



2.0 **PRODUCT DESCRIPTION**

Product Code:	Multi Analyte Custom Grade Solution	
Catalog Number:	IV-STOCK-77	
Lot Number:	U2-MEB735844	
Matrix:	1% (v/v) HNO3	
Value / Analyte(s):	1 μg/L ea:	
	Beryllium,	Cerium,
	Iron,	Indium,
	Lithium,	Magnesium,
	Lead,	Uranium

3.0 **CERTIFIED VALUES AND UNCERTAINTIES**

ANALYTE Beryllium, Be	CERTIFIED VALUE 1.000 ± 0.007 μg/L	ANALYTE Cerium, Ce	CERTIFIED VALUE 1.000 ± 0.007 µg/L
Indium, In	1.000 ± 0.006 μg/L	Iron, Fe	1.000 ± 0.007 μg/L
Lead, Pb	1.000 ± 0.006 μg/L	Lithium, Li	1.000 ± 0.008 μg/L
Magnesium, Mg	1.000 ± 0.007 μg/L	Uranium, U	1.000 ± 0.007 μg/L

Density:

1.004 g/mL (measured at 20 ± 4 °C)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
Be	Calculated		See Sec. 4.2
Ce	Calculated		See Sec. 4.2
Fe	Calculated		See Sec. 4.2
In	Calculated		See Sec. 4.2
Li	Calculated		See Sec. 4.2
Mg	Calculated		See Sec. 4.2
Pb	ICP Assay	3128	101026
Pb	Calculated		See Sec. 4.2
U	Calculated		See Sec. 4.2

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods Certified Value, X _{CRM/RM} , where two or more methods of characterization are used is the weighted mean of the results:	Characterization of CRM/RM by One Method Certified Value, X _{CRM/RM} , where one method of characterization is used is the mean of individual results:
X _{CRM/RM} = Σ(w _i) (X _i) X _i = mean of Assay Method i with standard uncertainty u _{char} i w _i = the weighting factors for each method calculated using the inverse square of the variance: w _i = (1/u _{char}) ² / (Σ(1/(u _{char}) ²))	$X_{CRM/RM} = (X_a) (u_{char a})$ $X_a = mean of Assay Method A withu_{char a} = the standard uncertainty of characterization Method A$
CRM/RM Expanded Uncertainty (±) = U _{CRM/RM} = k $(u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{\frac{1}{2}}$ k = coverage factor = 2 $u_{char} = [\Sigma((w_i)^2 (u_{char}i)^2)]^{\frac{1}{2}}$ where $u_{char}i$ are the errors from each characterization method u_{bb} = bottle to bottle homogeneity standard uncertainty u_{ts} = long term stability standard uncertainty (storage) u_{ts} = transport stability standard uncertainty u_{ts} = transport stability standard uncertainty ertified Abundance:	$\begin{split} & \textbf{CRM/RM Expanded Uncertainty (±) = U_{CRM/RM} = k (u_{char a}^2 + u_{bb}^2 + u_{lts}^2 + u_{ts}^2)^{\frac{1}{2}}} \\ & \textbf{k} = \text{coverage factor = 2} \\ & \textbf{u}_{char a} = \text{the errors from characterization} \\ & \textbf{u}_{bb} = \text{bottle to bottle homogeneity standard uncertainty} \\ & \textbf{u}_{lts} = \text{long term stability standard uncertainty} \\ & \textbf{u}_{ts} = \text{transport stability standard uncertainty} \end{split}$

IV's Certified Abundance	
Isotope	Atom %
Uranium 238U	99.8 ± 0.1
Uranium 235U	0.19 ± 0.05

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

N/A

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM[™]) see the Limited License to Use PCRM[™] in the Inorganic Ventures <u>Terms and Conditions of Sale</u>. <u>https://www.inorganicventures.com/terms-and-conditions-sale</u>. The Terms and Conditions contain information on the use of materials traceable to PCRM[™] certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between $4^{\circ} - 24^{\circ}$ C to minimize the effects of transpiration. Use at $20^{\circ} \pm 4^{\circ}$ C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit www.inorganicventures.com/TCT

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

August 31, 2023

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- August 31, 2028

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

NAMES AND SIGNATURES OF CERTIFYING OFFICERS 12.0 **Certificate Prepared By:**

Uyen Truong **Custom Processing Supervisor**

Uyu m ming

Certificate Approved By:

Muzzammil Khan Stock Laboratory Supervisor

Mayn Kh Paul R Saines

Certifying Officer:

Paul Gaines Chairman / Senior Technical Director