

300 Technology Drive
 Christiansburg, VA 24073 USA
 inorganicventures.com

 P: 800-669-6799/540-585-3030
 F: 540-585-3012
 info@inorganicventures.com

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).



2.0 PRODUCT DESCRIPTION

Product Code: Single Analyte Mass Spec Solution
 Catalog Number: MSGE-100PPM
 Lot Number: V2-GE740222
 Matrix: tr. HNO₃
 tr. HF
 Value / Analyte(s): 100 µg/mL ea:
 Germanium
 Starting Material: Ge metal
 Starting Material Lot#: 2105
 Starting Material Purity: 99.9925%

3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Value: 100.06 ± 0.71 µg/mL
Density: 0.999 g/mL (measured at 20 ± 4 °C)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
Ge	ICP Assay	3120a	151115

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$$X_{CRM/RM} = \sum(w_i) (X_i)$$

X_i = mean of Assay Method i with standard uncertainty $u_{char i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$$w_i = (1/u_{char i}^2) / (\sum(1/u_{char i}^2))$$

$$CRM/RM \text{ Expanded Uncertainty } (\pm) = U_{CRM/RM} = k (u_{char}^2 + u_{bb}^2 + u_{lts}^2 + u_{ts}^2)^{1/2}$$

k = coverage factor = 2

u_{char} = $[\sum(w_i)^2 (u_{char i}^2)]^{1/2}$ where $u_{char i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$$X_{CRM/RM} = (X_a) (u_{char a})$$

X_a = mean of Assay Method A with

$u_{char a}$ = the standard uncertainty of characterization Method A

$$CRM/RM \text{ Expanded Uncertainty } (\pm) = U_{CRM/RM} = k (u_{char a}^2 + u_{bb}^2 + u_{lts}^2 + u_{ts}^2)^{1/2}$$

k = coverage factor = 2

$u_{char a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

CRM/RMs are tested for trace metallic impurities by Axial ICP-OES and ICP-MS. The result from the most sensitive method for each element, is reported below. Solutions tested by ICP-MS were analyzed in an ULPA-Filtered Clean Room. An ULPA-Filter is 99.9985% efficient for the removal of particles down to 0.3 µm.

M	Ag	<	0.001601	M	Eu	<	0.000200	O	Na		0.000036	M	Se	<	0.018620	O	Zn		0.000012
O	Al	<	0.011554	O	Fe		0.000032	O	Nb	<	0.005555	O	Si		0.000123	O	Zr	<	0.004444
O	As	<	0.006666	M	Ga	<	0.001601	M	Nd	<	0.000200	M	Sm	<	0.001601				
M	Au	<	0.003407	M	Gd	<	0.000200	M	Ni	<	0.002803	M	Sn		0.000017				
O	B	<	0.016665	s	Ge	<		M	Os	<	0.000200	O	Sr	<	0.000666				
O	Ba		0.000067	M	Hf		0.000009	O	P	<	0.022220	O	Ta	<	0.022220				
O	Be	<	0.000333	M	Hg	<	0.000400	M	Pb	<	0.000200	M	Tb	<	0.000200				
M	Bi	<	0.001601	M	Ho	<	0.000200	M	Pd	<	0.003207	M	Te	<	0.004404				
O	Ca		0.000055	M	In	<	0.016017	M	Pr	<	0.000200	M	Th	<	0.002803				
O	Cd	<	0.001555	M	Ir	<	0.000200	M	Pt	<	0.000200	O	Ti		0.000086				
M	Ce	<	0.000200	O	K		0.000036	M	Rb	<	0.000800	M	Tl		0.000023				
M	Co	<	0.000200	M	La	<	0.000200	M	Re	<	0.000200	M	Tm	<	0.000200				
M	Cr	<	0.001001	O	Li	<	0.000022	M	Rh	<	0.000200	M	U	<	0.000200				
M	Cs	<	0.001201	M	Lu	<	0.000200	M	Ru	<	0.000801	M	V	<	0.001201				
M	Cu	<	0.000600	O	Mg		0.000005	O	S		0.007642	M	W		0.000025				
M	Dy	<	0.000200	M	Mn	<	0.000200	M	Sb		0.000153	O	Y	<	0.000555				
M	Er	<	0.000200	M	Mo	<	0.005205	O	Sc	<	0.000666	M	Yb	<	0.000200				

M - Checked by ICP-MS O - Checked by ICP-OES i - Spectral Interference
n - Not Checked For s - Solution Standard Element

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures [Terms and Conditions of Sale](https://www.inorganicventures.com/terms-and-conditions-sale). <https://www.inorganicventures.com/terms-and-conditions-sale>. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit www.inorganicventures.com/TCT

Atomic Weight; Valence; Coordination Number; Chemical Form in Solution - 72.59 +4 6 Ge(OH)x(F)y2-

Chemical Compatibility -Stable in HCl, HF, H3PO4 H2SO4 and HNO3 as the Ge(OH)x(F)y2-. Avoid neutral to basic media. Unstable at ppm levels with metals that would pull F- away (i.e. Do not mix with Alkaline or Rare Earths or high levels of transition elements unless they are fluorinated). Stable with most inorganic anions with a tendency to hydrolyze .

Stability - 2-100 ppb levels - stability unknown alone or mixed with all other metals as the Ge(OH)x(F)y2-. 1-10,000 ppm single element solutions as the Ge(OH)x(F)y2- chemically stable for years in 2-5% HNO3 / trace HF in a LDPE container.

Ge Containing Samples (Preparation and Solution) - Metal (Soluble in 1:1:1 H2O / HF / HNO3); Oxide - GeO (Readily soluble in HCl or NaOH), GeO2 (fuse in Pt0with Na2CO3 followed by HCl solution of the fuseate); Geological Samples (fuse in Pt0with Na2CO3 followed by HCl solution of the fuseate); Organic Matrices (Dry ash at 450EC in Pt0 and dissolve by gently warming with 1:1:1 H2O / HF / H2SO4 or fuse ash with Na2CO3 and dissolve fuseate with HCl / H2O).

Atomic Spectroscopic Information (ICP-OES D.L.s are given as radial/axial view):

Technique/Line	Estimated D.L.	Order	Interferences (underlined indicates severe)
ICP-MS 72 amu	20	n/a	36Ar2,37Cl17O18O, 37Cl35Cl,36S18O2, 36S2,36Ar36S,56Fe 16O,40Ar16O2,40C a16O2,40Ar32S,144 Nd2+,44Sm2+
ICP-OES 164.919 nm	0.01 / 0.001 µg/mL	1	Co, Fe, Cu
ICP-OES 219.871 nm	0.06 / 0.009 micro:g/mL	1	W, Ir, Re, Co
ICP-OES 265.117 nm	0.05 / 0.009 micro:g/mL	1	Ta, Pt, Mn, Rh, Ce, Nb, Hf

HF Note: This standard should not be prepared or stored in glass.

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

January 31, 2024

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- **January 31, 2029**

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS


Certificate Prepared By:

Uyen Truong
Custom Processing Supervisor



Certificate Approved By:

Thomas Kozikowski
Stock VS Manager



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director

