

Certificate of Analysis

300 Technology Drive Christiansburg, VA 24073 USA inorganicventures.com P: 800-669-6799/540-585-3030 F: 540-585-3012 info@inorganicventures.com

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).



2.0 PRODUCT DESCRIPTION

Product Code:	Single Analyte Mass Spec Solution				
Catalog Number:	MSSN-10PPM				
Lot Number:	U2-SN734521				
Matrix:	5% (v/v) HNO3 tr. HF				
Value / Analyte(s):	10 μg/mL ea: Tin				
Starting Material:	Tin Metal				
Starting Material Lot#:	1963				
Starting Material Purity:	99.9962%				
CERTIFIED VALUES AND UNCERTAINTIES					

Certified Value:	10.008 ± 0.075 μg/mL
Density:	1.025 g/mL (measured at 20 ± 4 °C)

Assay Information:

3.0

ANALYTE	METHOD	NIST SRM#	SRM LOT#
Sn	ICP Assay	3161a	140917
Sn	Calculated		See Sec. 4.2

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods	Characterization of CRM/RM by One Method
Certified Value, X _{CRM/RM} , where two or more methods of characterization are used is the weighted mean of the results:	Certified Value, X _{CRWRM} , where one method of characterization is used is the mean of individual results:
X _{CRM/RM} = Σ(w _i) (X _i) X _i = mean of Assay Method i with standard uncertainty u _{char} i w _i = the weighting factors for each method calculated using the inverse square of the variance: w _i = (1/u _{char}) ² / (Σ(1/(u _{char}) ²)	$X_{CRM/RM} = (X_a) (u_{char a})$ $X_a =$ mean of Assay Method A with $u_{char a}$ = the standard uncertainty of characterization Method A
CRM/RM Expanded Uncertainty (±) = U _{CRM/RM} = k ($u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts}$) ^{1/2} k = coverage factor = 2 $u_{char} = [\Sigma((w_i)^2 (u_{char}_i)^2)]^{1/2}$ where u_{char}_i are the errors from each characterization method u_{bb} = bottle to bottle homogeneity standard uncertainty u_{lts} = long term stability standard uncertainty (storage) u_{ts} = transport stability standard uncertainty	$\begin{split} & CRM/RM \text{ Expanded Uncertainty } (\pm) = U_{CRM/RM} = k \left(u^2_{\ char \ a} + u^2_{\ bb} + u^2_{\ lts} + u^2_{\ ts} \right)^{1/2} \\ & k = coverage factor = 2 \\ & u_{char \ a} = the errors from characterization \\ & u_{bb} = bottle to bottle homogeneity standard uncertainty \\ & u_{lts} = long term stability standard uncertainty (storage) \\ & u_{ts} = transport stability standard uncertainty \end{split}$

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

CRM/RMs are tested for trace metallic impurities by Axial ICP-OES and ICP-MS. The result from the most sensitive method for each element, is reported below. Solutions tested by ICP-MS were analyzed in an ULPA-Filtered Clean Room. An ULPA-Filter is 99.9985% efficient for the removal of particles down to 0.3 µm.

М	Ag		0.000003	м	Eu	<	0.002100	0	Na		0.000002	м	Se	<	0.018000	0	Zn		0.000001
0	AI	<	0.016000		Fe		0.000010		Nb	<	0.001100		Si		0.000010		Zr	<	0.011000
M	As	-	0.000007		Ga	<	0.000530		Nd	<	0.000530		Sm	/	0.000790	Ŭ	-	-	0.011000
															0.000790				
М	Au	<	0.000260	М	Gd	<	0.003200	0	Ni	<	0.001300	s	Sn	<					
0	В	<	0.005800	М	Ge	<	0.004800	М	Os	<	0.000260	М	Sr	<	0.000270				
0	Ва	<	0.001200	М	Hf		0.000001	0	Р	<	0.053000	М	Та		0.000004				
0	Be	<	0.000630	Μ	Hg	<	0.001100	Μ	Pb		0.000043	М	Tb	<	0.000530				
М	Bi		0.000047	Μ	Ho	<	0.000270	Μ	Pd	<	0.000780	0	Te	<	0.024000				
0	Ca		0.000006	0	In		0.000172	М	Pr	<	0.008400	М	Th	<	0.011000				
0	Cd		0.000001	М	lr	<	0.000260	М	Pt	<	0.000270	0	Ti		0.000062				
М	Ce	<	0.027000	0	К		0.000006	М	Rb	<	0.002100	М	ΤI	<	0.000270				
0	Co	<	0.002000	0	La	<	0.005900	М	Re	<	0.001400	М	Tm	<	0.000270				
М	Cr	<	0.001600	0	Li		0.000001	М	Rh	<	0.000530	М	U	<	0.027000				
М	Cs	<	0.040000	М	Lu	<	0.000270	М	Ru	<	0.001600	М	V	<	0.001100				
М	Cu		0.000002	0	Mg		0.000001	i	S	<		М	W		0.000001				
М	Dy	<	0.001600	0	Mn	<	0.000530	0	Sb		0.000008	М	Υ	<	0.000270				
М	Er	<	0.000270	М	Мо	<	0.001600	0	Sc	<	0.001200	М	Yb	<	0.000270				

M - Checked by ICP-MS O - Checked by ICP-OES i - Spectral Interference n - Not Checked For s - Solution Standard Element

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM[™]) see the Limited License to Use PCRM[™] in the Inorganic Ventures <u>Terms and Conditions of Sale</u>, <u>https://www.inorganicventures.com/terms-and-conditions-sale</u>. The Terms and Conditions contain information on the use of materials traceable to PCRM[™] certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between $4^{\circ} - 24^{\circ}$ C to minimize the effects of transpiration. Use at $20^{\circ} \pm 4^{\circ}$ C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit www.inorganicventures.com/TCT

Atomic Weight; Valence; Coordination Number; Chemical Form in Solution - 118.71 +4 4,5, 6,7,8 Sn(OH)xFy2-

Chemical Compatibility -Soluble in HCI and dilute HF / HNO3. Avoid neutral to basic media. Unstable at ppm levels with metals that would pull F- away. (i.e. Do not mix with Alkaline or Rare Earths or high levels of transition elements unless they are fluorinated.) Stable with most inorganic anions provided it is in the chemical form shown above.

Stability - 2-100 ppb levels stable (alone or mixed with all other metals that are at comparable levels) as the Sn(OH)xFy2- for 1 year in 1% HNO3 / LDPE container. 1-10,000 ppm single element solutions as the Sn(OH)xFy2- chemically stable for years in 2-5% HNO3 / trace HF in a LDPE container.

Sn Containing Samples (Preparation and Solution) - Metal (Soluble in HF / HNO3 or HCI); Oxides - SnO (soluble in HCI), SnO2 -very resistant to all acids including HF(Fusion with equal parts of Na2CO3 and S. It is then soluble in water or dilute acids as the thiostannate.); Alloys (Treat first 0.1 g with 10 mL conc. H2SO4 to boiling until the alloy disintegrates and nearly all of the sulfuric acid is expelled. Then add 100 mL O2 free water and 50 mL of conc HCl or transfer to a plastic container and add 1 mL HF in either case warming gently to bring about solution.); Organic Matrices (Volatility and precipitation of the insoluble stannic oxide are problems. Consultation of the literature should be made for individual matrices / Sn compounds.)

Atomic Spectroscopic Information (ICP-OES D.L.s are given as radial/axial view):

Technique/Line	Estimated D.L.	Order	Interferences (underlined indicates severe)
ICP-MS 120 amu	5 ppt	N/A	120Te, 104Ru16O,
			104Pd16O
ICP-OES 189.989 nm	0.03 / 0.003 µg/mL	1	
ICP-OES 242.949 nm	0.1 / 0.01 µg/mL	1	W, Mo, Rh ,Ta, Co

HF Note: This standard should not be prepared or stored in glass.

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.con

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

July 12, 2023

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- July 12, 2028
- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date:

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Approved By:

Thomas Kozikowski Stock VS Manager

SD9784. Paul R Saine

Certifying Officer:

Paul Gaines Chairman / Senior Technical Director