

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

ACCREDITATION / REGISTRATION

Inorganic Ventures is an ISO 9001 registered manufacturer (Amtivo Certificate Number 274357).

DISCRETE ANALYZER REAGENT Potassium antimonyl tartrate / Ammonium molybdate / Sulfuric acid solution

KEEP PRODUCT REFRIGERATED. See Storage and Handling Information.

Catalog No: RGT-PHOSPHATE-PO4-BTL-1-OF-2
Lot No: X2-DA763801

Starting Material	Potassium antimonyl tartrate trihydrate
Starting Material Lot No:	IV-2585

Starting Material:	Ammonium molybdate tetrahydrate
Starting Material Lot No:	IV-3003

Starting Material:	Sulfuric acid
Starting Material Lot No:	2512662012

NOMINAL CONCENTRATION: 0.216 g/L Potassium antimonyl tartrate trihydrate / 8.56 g/L
Ammonium molybdate tetrahydrate / 3.57 N Sulfuric acid

This solution is a reagent and is not intended to be used as a certified reference material.

PREPARATION

This reagent solution is ready to use without further preparation for analyzing o-Phosphate (PO_4) in water utilizing colorimetric methods based on method chemistry described in EPA Method 365.1 Rev 2.0.

INSTRUCTIONS FOR THE CORRECT USE OF THIS REAGENT

STORAGE AND HANDLING

Store refrigerated at $4^{\circ} \pm 2^{\circ}$ C while in sealed TCT bag. After opening the sealed TCT bag keep cap tightly sealed when not in use. Continue to store bottle at $4^{\circ} \pm 2^{\circ}$ C when not in use. Allow for reagent to reach room temperature and shake well before use.

Loss of method sensitivity and/or reagent discoloration has been observed when not following the recommended storage conditions.

BEST PRACTICES

Do not return removed aliquots of the reagent back to the container. Do not mix different lot numbers of the same reagent during sample analysis. A fresh calibration should be performed after refilling reagents.

This reagent must be pipetted within $\pm 5\%$ of the target volume to ensure an accurate analysis. The use of a calibrated pipette is recommended. Validate pipetting technique to ensure tolerance within this range.

Certification Date: 08/19/2026

Prepared By:



Approved By:



Expiration Date: 08/19/2028

Certifying Officer:

