



Refine your results. Redefine your industry.

Certificate of Analysis

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

ACCREDITATION / REGISTRATION

Inorganic Ventures is an ISO 9001 registered manufacturer (Amtivo Certificate Number 274357).

DISCRETE ANALYZER REAGENT Ascorbic acid solution

KEEP PRODUCT REFRIDGERATED. See Storage and Handling Information.

Catalog No: RGT-PHOSPHATE-PO4-BTL-2-OF-2
Lot No: X2-DA763193

Starting Material:	Ascorbic acid
Starting Material Lot No:	IV-2884

NOMINAL CONCENTRATION: 18 g/L Ascorbic acid

This solution is a reagent and is not intended to be used as a certified reference material.

PREPARATION

This reagent solution is ready to use without further preparation for analyzing Phosphate (PO_4) in water utilizing colorimetric methods based on method chemistry described in EPA Method 365.1 Rev 2.0.

INSTRUCTIONS FOR THE CORRECT USE OF THIS REAGENT

STORAGE AND HANDLING

Store refrigerated at $4^\circ\text{C} \pm 2^\circ\text{C}$ while in sealed TCT bag. After opening the sealed TCT bag keep cap tightly sealed when not in use. Continue to store bottle at $4^\circ\text{C} \pm 2^\circ\text{C}$ when not in use. Do not return removed aliquots to container. Allow for reagent to reach room temperature and shake well before use.

Loss of method sensitivity and/or reagent discoloration has been observed when not following the recommended storage conditions.

BEST PRACTICES

Do not return removed aliquots of the reagent back to the container. Do not mix different lot numbers of the same reagent during sample analysis. A fresh calibration should be performed after refilling reagents.

This reagent must be pipetted within $\pm 5\%$ of the target volume to ensure an accurate analysis. The use of a calibrated pipette is recommended. Validate pipetting technique to ensure tolerance within this range.

Certification Date: 06/12/2026

Prepared By:



Expiration Date: 06/12/2028

Approved By:



Certifying Officer:

