

CERTIFICATE OF ANALYSIS

# 1.0 ACCREDITATION / REGISTRATION

**INORGANIC VENTURES** is accredited to ISO Guide 34, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (SAI Global File Number 010105).

# 2.0 PRODUCT DESCRIPTION

Product Code:	Single Analyte Atomic Absorption Solution
Catalog Number:	AAMG1
Lot Number:	G2-MG03121
Matrix:	0.1% HNO3 (v/v)
Value/Analyte(s):	1 000 µg/mL Magnesium

# 3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Value:	1000 ± 10 μg/mL
Certified Density:	1.003 g/mL (measured at 20 ± 1 °C)

## 4.0 TRACEABILITY TO NIST

The concentration of this solution standard has been verified by Inductively Coupled Plasma Spectroscopy (ICP) and is traceable to NIST SRM 3131a

## 4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

## 4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

## 4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

# 5.0 TRACE METALLIC IMPURITIES (TMI ) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

N/A

## 6.0 INTENDED USE

- For the calibration of analytical instruments and validation of analytical methods as appropriate.

## 7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

## 7.1 Storage and Handling Recommendations

- Keep cap tightly sealed when not in use. Store and use at  $20 \pm 4^{\circ}$  C. Do not pipette from the container. Do not return removed aliquots to container.

**Chemical Compatibility** -Soluble in HCl, HNO3, and H2SO4 avoid HF, H3PO4 and neutral to basic media. Stable with most metals and inorganic anions forming insoluble silicates, carbonates, hydroxides, oxides, and tungstates in neutral and slightly acidic media.

**Stability** - 2-100 ppb levels stable for months in 1% HNO3 / LDPE container. 1-10,000 ppm solutions chemically stable for years in 1-10% HNO3 / LDPE container.

**Mg Containing Samples (Preparation and Solution)** -Metal (Best dissolved in diluted HNO3); Oxide (Readily soluble in above compatible aqueous acidic solutions); Ores (Carbonate fusion in Pt0 followed by HCI dissolution); Organic Matrices (Sulfuric / peroxide digestion or nitric / sulfuric / perchloric acid decomposition, or dry ash and dissolution in dilute HCI).

# Atomic Spectroscopic Information (ICP-OES D.L.s are given as radial/axial view):

Technique/Line	Estimated D.L.	Order	Interferences (underlined indicates severe
ICP-MS 24 amu	42 ppt	n/a	7Li17O, 48Ti+2 ,
			48Ca+2
ICP-OES 279.553 nm	0.0002 / 0.00003 µg/mL	1	Th
ICP-OES 280.270 nm	0.0003 / 0.00005 μg/mL	1	U, V
ICP-OES 285.213 nm	0.002 / 0.00003 μg/mL	1	U, Hf, Cr, Zr

# 8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

# 9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

## 10.0 QUALITY STANDARD DOCUMENTATION

# 10.1 10CFR50 Appendix B - Nuclear Regulatory Commission

- Domestic Licensing of Production and Utilization Facilities

## 10.2 10CFR21 - Nuclear Regulatory Commission

- Reporting defects and Non-Compliance

## 10.3 ISO 9001 Quality Management System Registration

- SAI Global File Number 010105

# 11.0 CERTIFICATION, EXPIRATION AND PERIOD OF VALIDITY

## 11.1 Certification Issue Date

June 03, 2013

## 11.2 Expiration Date

## 11.3 Period of Validity

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is handled and stored in accordance with instructions given in Sec 7.0 and used prior to the date given in Sec 11.2. This certification is nullified if the CRM/RM is damaged, contaminated, or otherwise modified.

#### 12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

**Certificate Prepared By:** 

Donna Senn Product Documentation Technician

Donne Sena Elizabet a. Day Paul R. Lainea

## **Certificate Approved By:**

Elizabeth Day Quality Assurance Specialist

## **Certifying Officer:**

Paul Gaines PhD., Senior Technical Director