

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO Guide 34, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (SAI Global File Number 010105).



2.0 PRODUCT DESCRIPTION

Product Code: Multi Analyte Custom Grade Solution
 Catalog Number: SEP-862
 Lot Number: J2-MEB582071
 Matrix: 5% (v/v) HCl
 Value / Analyte(s): 10 000 µg/mL ea:
 Zr

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ANALYTE	CERTIFIED VALUE	ANALYTE	CERTIFIED VALUE
Zirconium, Zr	10 126.0 ± 56.0 µg/mL		

Certified Density: 1.028 g/mL (measured at 20 ± 1 °C)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
Zr	ICP Assay	3169	071226

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

$$\text{Certified Value } (\bar{x}) = \frac{\sum x_i}{n}$$

(\bar{x}) = mean
 x_i = individual results
 n = number of measurements

$$\text{Uncertainty } (\pm) = 2 \left[\sum (s_i)^2 \right]^{1/2}$$

2 = the coverage factor.
 $\left[\sum (s_i)^2 \right]^{1/2}$ = The square root of the sum of the squares of the most common errors (where 's' stands for the standard deviation) from instrumental measurement, density, NIST SRM uncertainty, weighing, dilution to volume, homogeneity, long term stability and short term stability.

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

CRM/RMs are tested for trace metallic impurities by Axial ICP-OES and ICP-MS. The result from the most sensitive method for each element, is reported below. Solutions tested by ICP-MS were analyzed in an ULPA-Filtered Clean Room. An ULPA-Filter is 99.9985% efficient for the removal of particles down to 0.3 µm.

i	Ag	<		M	Eu	<	0.000100	O	Na		0.018420	M	Se	<	0.020017	O	Zn		0.044004
M	Al		0.011469	O	Fe		0.008084	M	Nb	<	0.040033	O	Si		0.020467	s	Zr	<	
i	As	<		M	Ga	<	0.004003	M	Nd	<	0.000400	M	Sm	<	0.001201				
M	Au	<	0.005021	M	Gd		0.007480	M	Ni	<	0.001601	M	Sn	<	0.001401				
M	B	<	0.002202	M	Ge	<	0.002002	M	Os	<	0.000100	M	Sr		0.031713				
M	Ba		0.004887	M	Hf		0.162655	O	P	<	0.030810	M	Ta	<	0.001001				
M	Be	<	0.000801	M	Hg	<	0.000201	M	Pb		0.001695	M	Tb	<	0.000100				
M	Bi		0.009175	M	Ho	<	0.000100	M	Pd	<	0.050042	M	Te	<	0.008207				
O	Ca		0.017397	M	In	<	0.000400	M	Pr	<	0.000100	M	Th	<	0.000100				
O	Cd	<	0.007189	M	Ir	<	0.000201	M	Pt	<	0.006005	O	Ti	<	0.030810				
M	Ce		0.003690	O	K		0.050144	M	Rb	<	0.000801	M	Tl	<	0.001001				
M	Co	<	0.001401	M	La	<	0.000100	M	Re	<	0.000100	M	Tm	<	0.000400				
O	Cr	<	0.020540	M	Li	<	0.001501	M	Rh	<	0.000801	M	U	<	0.000100				
M	Cs	<	0.001201	M	Lu	<	0.002002	M	Ru	<	0.000100	O	V	<	0.030810				
M	Cu		0.016056	O	Mg		0.003275	O	S	<	0.020540	M	W	<	0.008207				
M	Dy	<	0.000200	M	Mn		0.003092	M	Sb	<	0.011009	M	Y	<	0.000400				
M	Er	<	0.000200	M	Mo	<	0.001101	O	Sc	<	0.004108	M	Yb	<	0.000100				

M - Checked by ICP-MS O - Checked by ICP-OES i - Spectral Interference
n - Not Checked For s - Solution Standard Element

6.0 INTENDED USE

- For the calibration of analytical instruments and validation of analytical methods as appropriate.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Keep cap tightly sealed when not in use. Store and use at 20 ± 4° C. Do not pipette from the container. Do not return removed aliquots to container.

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 10CFR50 Appendix B - Nuclear Regulatory Commission

- Domestic Licensing of Production and Utilization Facilities

10.2 10CFR21 - Nuclear Regulatory Commission

- Reporting defects and Non-Compliance

10.3 ISO 9001 Quality Management System Registration

- SAI Global File Number 010105

10.4 ISO/IEC Guide 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.5 ISO/IEC Guide 34 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

11.0 CERTIFICATION, EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

June 03, 2015

11.2 Expiration Date


11.3 Period of Validity

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is handled and stored in accordance with instructions given in Sec 7.0 and used prior to the date given in Sec 11.2. This certification is nullified if the CRM/RM is damaged, contaminated, or otherwise modified.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS


Certificate Prepared By:

Donna Senn
Product Documentation Technician



Certificate Approved By:

Brian Alexander
PhD., Technical Process Director



Certifying Officer:

Paul Gaines
PhD., Senior Technical Director

